

Original citation:

Dale, M. W. and Morley, Gavin (2017) *Medical applications of diamond magnetometry : commercial viability*. Discussion Paper. arXiv. 1705.01994 (Unpublished)

Permanent WRAP URL:

<http://wrap.warwick.ac.uk/100522>

Copyright and reuse:

The Warwick Research Archive Portal (WRAP) makes this work by researchers of the University of Warwick available open access under the following conditions. Copyright © and all moral rights to the version of the paper presented here belong to the individual author(s) and/or other copyright owners. To the extent reasonable and practicable the material made available in WRAP has been checked for eligibility before being made available.

Copies of full items can be used for personal research or study, educational, or not-for-profit purposes without prior permission or charge. Provided that the authors, title and full bibliographic details are credited, a hyperlink and/or URL is given for the original metadata page and the content is not changed in any way.

A note on versions:

The version presented here is a working paper or pre-print that may be later published elsewhere. If a published version is known of, the above WRAP URL will contain details on finding it.

For more information, please contact the WRAP Team at: wrap@warwick.ac.uk

Medical applications of diamond magnetometry: commercial viability

Matthew W. Dale and Gavin W. Morley

Department of Physics, University of Warwick

May 8, 2017

The sensing of magnetic fields has important applications in medicine, particularly to the sensing of signals in the heart and brain. The fields associated with biomagnetism are exceptionally weak, being many orders of magnitude smaller than the Earth's magnetic field. To measure them requires that we use the most sensitive detection techniques, however, to be commercially viable this must be done at an affordable cost. The current state of the art uses costly SQUID magnetometers, although they will likely be superseded by less costly, but otherwise limited, alkali vapour magnetometers. Here, we discuss the application of diamond magnetometers to medical applications. Diamond magnetometers are robust, solid state devices that work in a broad range of environments, with the potential for sensitivity comparable to the leading technologies.

1 Introduction

A magnetometer is a device that senses a magnetic field, measuring its strength and sometimes direction too. There are many different applications of magnetometry such as navigation and geo-surveying, but the one we focus on here is the measurement of medical biomagnetism, and specifically the magnetic fields produced by the heart.

1.1 Magnetocardiography

The cardiac cycle is initiated by an electrical impulse from the sinoatrial node, which then propagates through the heart. The electrical activity passes through specialised pathways, controlling the timing of the contraction of the tissue. The measurement of magnetic fields created by the heart is called magnetocardiography (MCG), and provides

similar information to electrocardiography (ECG) which is widely used in hospitals. Several reviews of MCG are available, demonstrating that this technique provides diagnostic information additional to ECG for coronary artery disease and cardiac arrhythmias [1, 2, 3]. The MCG that has been carried out to date has used SQUID (superconducting quantum interference device) sensors at cryogenic temperatures. For this it is necessary to record magnetic fields with a sensitivity of $50 \text{ fT Hz}^{-1/2}$ or better with a temporal resolution of around 10 ms. The frequency range from DC to 50 Hz is the most useful diagnostically, and frequencies from 50 Hz to 250 Hz are also useful. Existing SQUID-based MCG achieves sensitivity of better than $10 \text{ fT Hz}^{-1/2}$ in a shielded room, or $40 \text{ fT Hz}^{-1/2}$ in an unshielded room [2]. The installation of magnetically shielded rooms is expensive (up to \$1M with multiple shielding layers) and inconvenient for hospitals and the need for this would hamper attempts to bring MCG into common use. Existing SQUID-based shielded MCG systems cost around \$1M, but the next generation using diamond or alkali vapour cells should sell for significantly less due to the partial or total elimination of shielding and the costs related to cryogenics. Currently the non-industrialised cost for making SQUID sensors is about \$1k. The cost of new magnetic sensors will have to be competitive with SQUID sensors following industrialisation of the manufacturing processes. Detecting the smaller magnetic fields from human brains is also of interest, and this is called magnetoencephalography (MEG).

1.2 Underlying physics of the negative nitrogen-vacancy centre in diamond

Diamond is a remarkable material being best in class for many properties; it has excellent thermal conductivity, hardness and wide transparency to electromagnetic radiation to name a few. In recent years another application of diamond has emerged and that is as a sensor, owing to some unique properties of impurities that can be incorporated into the carbon lattice.

The band-gap of diamond is 5.5 eV, making it transparent to all visible light and into the UV, however impurities, or colour centres as they are called, introduce energy levels into the band-gap. Transitions involving these energy levels allow absorption and emission of visible light giving a diamond its colour. The most common colour centre in diamond is substitutional nitrogen, as it is abundant in the atmosphere and easily fits into the diamond lattice. Another is the vacancy, which may be incorporated during crystal growth as an imperfection, or introduced post growth by irradiation with high energy particles. If a vacancy becomes trapped adjacent to a substitutional nitrogen the nitrogen vacancy centre (NV) is formed, a model of which is shown in Figure 1. A substitutional nitrogen centre can donate an electron to NV to make it negatively charged (NV^-).

It is the NV^- centre which has some special properties and as such has been the subject of a considerable body of research. It is photostable and exhibits red photoluminescence that can be detected from a single centre [4]. The ground state has an unpaired electron spin that can be polarised optically with very high efficiency [5]. The same mechanism that is responsible for the polarization also reduces the luminescence intensity of the unpolarised states by up to 30 %. These properties allow the optical detection of the




Figure 1: The nitrogen vacancy centre in diamond in its four orientations in the crystal lattice. The nitrogen atom is blue and the three carbon atoms surrounding the vacancy have been highlighted.

spin of a single NV^- centre and consequently optically detected magnetic resonance (ODMR). An excellent review is given by Doherty *et al.* of the physics of the NV^- centre [6].

A magnetic field splits the electron spin states, moving the position of the electron spin resonances, which can be read by ODMR. The accuracy of the field measurement is partly determined by the width of the resonance, which is primarily determined by the spin coherence lifetime. Diamond exhibits the longest coherence at room temperature of any solid state system [7, 8].

The long coherence times stem from several properties of diamond. Firstly, the spin-lattice relaxation is small because of the weak spin-orbit interaction and strong covalent bonding of diamond. Secondly, decoherence due to interaction with other spins is minimal. At natural isotopic abundance diamond is 98.9 % nuclear-spin free ^{12}C . In addition, there are few electron spins other than the NV^- centres we are interested in. The coherence times can be improved by engineering the diamond to remove parasitic defects [9], and enrichment with ^{12}C [8].

The NV^- centre has trigonal symmetry meaning that there are four possible orientations in which the N-V axis can sit in the diamond lattice; these are shown relative to each other in Figure 1. The resonance frequency depends not only on the magnitude of the magnetic field, but also on the angle between the field and the N-V axis. Measuring the resonances from three or four sites of the NV^- centre allows vector measurements of the magnetic field.

1.3 Nitrogen-vacancy sensitivity

A review of NV^- magnetometry in the period up to 2014 is available [10]. We should focus on demonstrated sensitivities rather than projected values. For bulk experiments, the best broadband sensitivity is $15 \text{ pT Hz}^{-1/2}$, from a diamond at room temperature with volume $13 \times 200 \times 2000 \text{ }\mu\text{m}^3$ for magnetic fields at frequencies from 80 Hz to 2000 Hz [11]. In a different experiment, using AC measurements instead of broadband (pulsed magnetic resonance instead of continuous-wave), the sensitivity to magnetic field at a particular frequency is higher, reaching $0.9 \text{ pT Hz}^{-1/2}$ for a 20 kHz magnetic field with a bulk diamond [12].

It has been proposed that using NV^- centres as a laser medium could provide significant further gains in sensitivity. By using the field dependence of the fluorescence to push the laser above threshold, contrast would be greatly increased. A so-called laser threshold magnetometer (LTM) based on NV^- is predicted to achieve a shot-noise limited DC sensitivity of $2 \text{ fT Hz}^{-1/2}$ [13].

1.4 Current intellectual property holders

There are a number of patents relating to diamond magnetometry, and in particular recently published by Lockheed Martin. Lockheed Martin patents cover vector magnetometry using ensembles of NV^- centres and a CW frequency swept experiment using a bias field to separate the spectrum [14]; a magnetometer using ensembles of NV^- cen-

tres with magnetic field detection by pulsed ODMR using an optimised Ramsey pulse sequence [15]; a method for determining the orientations of NV centres to calibrate the NV magnetometer [16]; using parabolic or ellipsoidal reflectors to improve collection efficiency of luminescence and increasing sensitivity by resolving nitrogen hyperfine resonances [17]; and micro sized NV magnetometers [18].

Sensitivity to magnetic field can be increased by making improvements to the diamond material. Element Six have a number of relevant patents in this area, in particular a patent targeted at producing diamond for spintronic applications [19]. The patent describes techniques for the growth of diamond by CVD with a low number of paramagnetic centres and ^{12}C isotopic enrichment.

A patent from Harvard describes improving the sensitivity of pulsed magnetometers by dynamical decoupling of electron spins from spin-spin interactions and interactions with the lattice [20].

2 Existing technologies

There is a large array of technologies for magnetic field detection. Presently the most sensitive and applicable to MCG or MEG are the well-established superconducting quantum interference devices (SQUIDs) and atomic vapour cells made with alkali metals. Each of these technologies will be discussed in more detail here along with their strengths and weaknesses.

2.1 SQUIDs

Since the first reports of the Josephson effect [21] and SQUID magnetometry over 50 years ago [22] they have been the subject of extensive research, driven in part by their potential application to MEG [23]. SQUIDs are one of the most sensitive magnetometers with typical sensitivities near $1\text{ fT Hz}^{-1/2}$ and a sensitivity of $0.5\text{ fT Hz}^{-1/2}$ at 1 kHz reported [24]. Current commercial MEG systems use arrays of SQUID magnetometers.

SQUIDs consist of a superconducting ring with one (RF SQUID) or more (DC SQUID) Josephson junctions. A small current less than a critical current can flow across the junction without creating a voltage. Because of flux quantisation, as the flux passing through the superconducting ring increases or decreases, the current flowing around it oscillates. When a bias current is applied this creates an oscillating voltage which can be read with conventional semiconductor electronics. The number of oscillations corresponds to the change in magnetic field.

The greatest limitation of SQUIDs is the requirement of cryogenic temperatures for operation. This increases cost and complexity and also reduces the sensitivity of a device since it must be some distance from the source of the fields being measured. In MEG the typical minimum distance between the SQUID sensor and brain is 30 mm [23]. The distance can be decreased to increase SNR at the expense of helium boil-off. Furthermore, in current designs the sensors and Dewar are fixed meaning it is not optimised to head size, reducing sensor proximity.

SQUIDs have been demonstrated using high T_C superconductors, with a sensitivity of $4 \text{ fT Hz}^{-1/2}$, allowing the use of nitrogen as a cryogen and closer sensor proximity [25], however there are significant challenges in their fabrication.

2.2 Alkali metal magnetometer

In the last decade the sensitivity of SQUID magnetometers has been challenged by atomic magnetometers (AM) using alkali metal vapours, which have the potential to replace SQUIDs in biomedical applications [26]. Such AMs detect the Faraday rotation [27] or absorption [28] of light through a spin polarised vapour of potassium, rubidium or caesium. Their operation does not require the cryogenic cooling with helium required by SQUIDs, significantly reducing the complexity and cost of a device.

The sensitivity of an alkali-metal magnetometer is limited by spin relaxation time, to which the dominant contribution is spin-exchange collisions. For high density vapours and very low magnetic fields, the atoms can exchange spin much faster than the magnetic precession frequency. This is known as the spin-exchange relaxation free (SERF) regime [29], and allows sub $\text{fT Hz}^{-1/2}$ sensitivities. Sensitivities such as $160 \text{ aT Hz}^{-1/2}$ with a measurement volume of 0.45 cm^3 have been reported [30].

3 Market

There are around 100 SQUID MEG systems installed worldwide, at a cost of over \$1M each. The MCG market should be much larger if the instrumentation was affordable and portable, because MCG has been shown to be superior to ECG and hence other non-invasive approaches for the diagnosis of coronary artery disease (CAD) [1, 2, 3]. CAD is the most common type of heart disease and is the leading cause of death in the United States in both men and women. Several companies have tried and failed to commercialize SQUID-based MCG, held back by the cost of a cryogen-based system. We estimate that 100,000 MCG systems could be sold if the functionality were the same as existing SQUID systems and the price was below \$150k. This is based on there being over 100,000 hospitals in China, India, the EU, Japan and the USA.

Diamond magnetometers are at technology readiness level (TRL) 7: the technology has been demonstrated and is moving towards being put on sale. However, this has not yet reached the sensitivity needed for MCG, so an MCG system based on diamond is at TRL 4-5 (technology development).

4 Outlook

Although established SQUID sensors have the necessary sensitivity for MCG and MEG, the requirement for cryogenic cooling with helium makes them ultimately too expensive for widespread commercial applications. In addition, the need for the sensors to be in a Dewar limits how close they can be positioned to the subject.

Alkali metal vapour magnetometers offer the significant advantage over SQUID magnetometers of removing the need for cryogenic cooling. They do require heating, but this is much less challenging and restrictive. The most sensitive alkali metal vapour cells, operating in the SERF regime, however require being in very low field environments necessitating the use of magnetically shielded rooms. The background magnetic noise must be significantly reduced for all highly sensitive magnetometers, and this can be achieved to a reasonable degree by subtracting the signal from additional sensors slightly removed from the subject, however a magnetically shielded environment is still a requirement for SERF magnetometers.

Whilst it is true that the sensitivity of diamond NV^- magnetometers is presently several orders of magnitude worse than the state of the art SQUID or alkali metal magnetometers, the aforementioned technologies have been in development for significantly longer. As such it is very likely that through optimising diamond material, sensor configuration, and detection methodology (for example a pulsed microwave experiment rather than frequency swept), orders of magnitude improvements can be made. Even if NV^- magnetometers do not ultimately exceed others in sensitivity they might offer significant advantages in ruggedness, cost, and proximity to the subject.

A large part of the cost for both alkali metal and NV^- magnetometers is the laser excitation. For alkali metal magnetometers, atomic transitions are excited, requiring finely tuned and stabilised lasers. The NV^- transition which is excited is significantly broadened by coupling to phonons in the lattice, effectively allowing broadband illumination. With LED technology constantly improving, single colour LEDs are now available with powers of hundreds of milliwatts. The use of LED illumination rather than laser has the potential to dramatically reduce the cost of a device.

A challenge with NV^- magnetometers is that the signal is contained in luminescence from the diamond. Luminescence radiates in all directions making its collection difficult (this is not a problem with alkali metal magnetometers since they detect laser light transmitted through the vapour cell). However the use of ellipsoidal reflectors is discussed in a Lockheed Martin patent [17], where it is stated that with correct implementation, near 100 % collection should be possible. A nice design would be the use of an ellipsoidal reflector coupled to a fibre, such that all light transmitted through the fibre is focussed on the diamond, and all luminescence from the diamond is coupled in to the fibre.

5 Acknowledgements

We have benefited from useful discussions with Peter Höfer (Bruker), Paul Mawson (Texas Instruments), Riccardo Fenici (BACPIC), Donatella Brisinda (BACPIC), Ian Fisher and Tim LeClair.

References

- [1] Joey S W Kwong, Boris Leithäuser, Jai-Wun Park, and Cheuk-Man Yu. Diagnostic value of magnetocardiography in coronary artery disease and cardiac arrhythmias:

- a review of clinical data. *International Journal of Cardiology*, 167(5):1835–42, 2013.
- [2] R. Fenici, D. Brisinda, A. Venuti, and A.R. Sorbo. Thirty years of clinical magnetocardiography at the Catholic University of Rome: Diagnostic value and new perspectives for the treatment of cardiac arrhythmias. *International Journal of Cardiology*, 168(5):5113–5115, 2013.
 - [3] Riccardo Fenici, Donatella Brisinda, and Anna Maria Meloni. Clinical application of magnetocardiography. *Expert Review of Molecular Diagnostics*, 5(3):291–313, may 2005.
 - [4] A Gruber, A Dräbenstedt, C Tietz, L Fleury, J Wrachtrup, and C von Borczyskowski. Scanning Confocal Optical Microscopy and Magnetic Resonance on Single Defect Centers. *Science*, 276(5321):2012–2014, 1997.
 - [5] M W Doherty, N B Manson, P Delaney, and L C L Hollenberg. The negatively charged nitrogen-vacancy centre in diamond: the electronic solution. *New Journal of Physics*, 13(2):025019, feb 2011.
 - [6] Marcus W. Doherty, Neil B. Manson, Paul Delaney, Fedor Jelezko, Jörg Wrachtrup, and Lloyd C.L. Hollenberg. The nitrogen-vacancy colour centre in diamond. *Physics Reports*, 528(1):1–45, jul 2013.
 - [7] Gopalakrishnan Balasubramanian, Philipp Neumann, Daniel Twitchen, Matthew Markham, Roman Kolesov, Norikazu Mizuochi, Junichi Isoya, Jocelyn Achard, Johannes Beck, Julia Tissler, Vincent Jacques, Philip R Hemmer, Fedor Jelezko, and Jörg Wrachtrup. Ultralong spin coherence time in isotopically engineered diamond. *Nature Materials*, 8(5):383–7, may 2009.
 - [8] T. Yamamoto, T. Umeda, K. Watanabe, S. Onoda, M. L. Markham, D. J. Twitchen, B. Naydenov, L. P. McGuinness, T. Teraji, S. Koizumi, F. Dolde, H. Fedder, J. Honert, J. Wrachtrup, T. Ohshima, F. Jelezko, and J. Isoya. Extending spin coherence times of diamond qubits by high-temperature annealing. *Physical Review B - Condensed Matter and Materials Physics*, 88:1–8, 2013.
 - [9] Y Chu, N P De Leon, B J Shields, B Hausmann, R Evans, E Togan, M J Burek, M Markham, A Stacey, A S Zibrov, A Yacoby, D J Twitchen, M Loncar, H Park, P Maletinsky, and M D Lukin. Coherent Optical Transitions in Implanted Nitrogen Vacancy Centers. *Nano letters*, 14:1982–1986, 2014.
 - [10] L Rondin, J P Tetienne, T Hingant, J F Roch, P Maletinsky, and V Jacques. Magnetometry with nitrogen-vacancy defects in diamond. *Rep Prog Phys*, 77(5):56503, 2014.
 - [11] John F. Barry, Matthew J. Turner, Jennifer M. Schloss, David R. Glenn, Yuyu Song, Mikhail D. Lukin, Hongkun Park, and Ronald L. Walsworth. Optical magnetic detection of single-neuron action potentials using quantum defects in diamond. *Proceedings of the National Academy of Sciences*, 113(49):14133–14138, 2016.

- [12] Thomas Wolf, Philipp Neumann, Kazuo Nakamura, Hitoshi Sumiya, Takeshi Ohshima, Junichi Isoya, and Jörg Wrachtrup. Subpicotesla diamond magnetometry. *Physical Review X*, 5(4):1–10, 2015.
- [13] Jan Jeske, Jared H. Cole, and Andrew D. Greentree. Laser threshold magnetometry. *New Journal of Physics*, 18(1):13015, 2016.
- [14] Peter G. Kaup and Arul Manickam. Apparatus and method for recovery of three dimensional magnetic field from a magnetic detection system, 2015.
- [15] Laird Nicholas Egan, David Nelson Coar, and Jon C. Russo. Apparatus and method for high sensitivity magnetometry measurement and signal processing in a magnetic detection system, 2016.
- [16] Arul Manickam, Peter G. Kaup, and John B. Stetson. Apparatus and method for estimating absolute axes’ orientations for a magnetic detection system, 2016.
- [17] Brian P. Boesch, Gregory Scott Bruce, Jeff D. Cammerata, David Nelson Coar, Laird Nicholas Egan, Bryan Neal Fisk, Wilbur Lew, Arul Manickam, Stephen Michael Sekelsky, John B. Stetson Jr, Peter G. Kaup, Julie Lynne Miller, Jon C. Russo, and Emanuel Solomon Stockman. Dnv magnetic field detector, 2016.
- [18] John B. Stetson Jr., Michael J. Hiltabidle, Peter G. Kaup, Arul Manickam, and Sarah Simon. Micro-dnv device, 2016.
- [19] Geoffrey Alan Scarsbrook, Daniel James Twitchen, and Matthew Lee Markham. Diamond material, 2010.
- [20] Mikhail D. Lukin, Ronald L. Walsworth, Amir Yacoby, Paola Cappellaro, Jacob M. Taylor, Liang Jiang, and Lilian Childress. Electronic spin based enhancement of magnetometer sensitivity, 2013.
- [21] P. W. Anderson and J. M. Rowell. Probable observation of the Josephson superconducting tunneling effect. *Physical Review Letters*, 10(6):230–232, 1963.
- [22] R. C. Jaklevic, John Lambe, A. H. Silver, and J. E. Mercereau. Quantum interference effects in Josephson tunneling. *Physical Review Letters*, 12(7):159–160, 1964.
- [23] Rainer Körber, Jan-Hendrik Storm, Hugh Seton, Jyrki P Mäkelä, Ritva Paetau, Lauri Parkkonen, Christoph Pfeiffer, Bushra Riaz, Justin F Schneiderman, Hui Dong, Seong-min Hwang, Lixing You, Ben Inglis, John Clarke, Michelle A Espy, Risto J Ilmonemi, Per E Magnelind, Andrei N Matlashov, Jaakko O Nieminen, Petr L Volegov, Koos C J Zevenhoven, Nora Höfner, Martin Burghoff, Keiji Enpuku, S Y Yang, Jen-Jei Chieh, Jukka Knuutila, Petteri Laine, and Jukka Nenonen. SQUIDS in biomagnetism: a roadmap towards improved healthcare. *Superconductor Science and Technology*, 29(11):113001, 2016.

- [24] T Fedele, H J Scheer, M Burghoff, G Curio, and R Körber. Ultra-low-noise EEG/MEG systems enable bimodal non-invasive detection of spike-like human somatosensory evoked responses at 1kHz. *Physiological measurement*, 36(2):357–68, 2015.
- [25] M. I. Faley, U. Poppe, R. E. Dunin-Borkowski, M. Schiek, F. Boers, H. Chocholacs, J. Dammers, E. Eich, N. J. Shah, A. B. Ermakov, V. Y. Slobodchikov, Y. V. Maslennikov, and V. P. Koshelets. High-Tc DC SQUIDs for magnetoencephalography. *IEEE Transactions on Applied Superconductivity*, 23(3), 2013.
- [26] Vishal K Shah and Ronald T Wakai. A compact, high performance atomic magnetometer for biomedical applications. *Physics in medicine and biology*, 58(22):8153–61, 2013.
- [27] D. Budker, D. Kimball, S. Rochester, V. Yashchuk, and M. Zolotarev. Sensitive magnetometry based on nonlinear magneto-optical rotation. *Physical Review A*, 62(4):1–7, 2000.
- [28] Vishal Shah, Svenja Knappe, Peter D. D. Schwindt, and John Kitching. Subpicotesla atomic magnetometry with a microfabricated vapour cell. *Nature Photonics*, 1(11):649–652, 2007.
- [29] J. C Allred, R. N Lyman, T. W Kornack, and M. V. Romalis. High-Sensitivity Atomic Magnetometer Unaffected by Spin-Exchange Relaxation. *Physical Review Letters*, 89(13):130801, 2002.
- [30] H. B. Dang, A. C. Maloof, and M. V. Romalis. Ultrahigh sensitivity magnetic field and magnetization measurements with an atomic magnetometer. *Applied Physics Letters*, 97(15):2014–2017, 2010.